


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *3377*  
Date: *Aug 25 '86*



- a) Output signal amplitude:  
*+25 TO +100 AMPS*
- b) Pulse width:  
*25 ± 5 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*~ 60 AMPS PIV + 15 ± 1.0 VOLT*  
*10 NSEC PIV 200 mA (MAX)*  
*100 KHZ*