


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *3378*  
Date: *SEPT 2 86*



- a) Output signal amplitude:  
*25 TO 100 AMPS*
- b) Pulse width:  
*25 ± 5 nSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 60 AMPS/DIV*  
*10 nSEC/DIV*  
*500 MHz*

*+15 ± 1.0 VOLT*  
*200 mA (MAX)*  
*[Signature]*