


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-AV-15 T*

S.N.: *3383*

Date: *NOV 24 86*



*Ⓐ 40 db ATTEN
500 KHz*



*Ⓑ 40 db ATTEN
500 KHz*

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 4.0 nsec
- c) Rise time:
≤ 60 psec
- d) Fall time:
≤ 135 psec
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+15 ± 1.0 VOLT
250 mA (MAX)*