


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-3-pw-c-pn-05*

S.N.: *3307*

Date: *NOV 28 86*



Ⓟ P OUT
 $R_L = 50 \Omega$
 50 VOLTS/DIV
 50 MHz



Ⓟ N OUT
 $R_L = 50 \Omega$
 50 VOLTS/DIV
 0.5 KHz

- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width:
0.1 TO 100 USEC.
- c) Rise time:
 ≤ 10 NSEC
- d) Fall time:
 ≤ 10 NSEC.
- e) PRF:
MIN PW: 0 TO 1 KHz
MAX PW: 0 TO 50 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 MHz
- h) OFFSET:
0 TO ± 50 VOLTS
0 TO ± 200 mA.