


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-1723-P*

S.N.: *3399*

Date: *NOV 20 '86*



*1.0 nSEC/DIV  
32 VOLTS/DIV  
(50 dB ATTEN)*

a) Output signal amplitude:  
*0 TO +60 VOLTS*

b) Pulse width:  
*≤ 1 nSEC*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 1.0 MHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*+15 ± 1.0 VOLTS  
200 mA (MAX)*