


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PVI-N-143-C-M1-F*

S.N.: *3400*

Date: *DEC 2 '86*



Ⓐ 60 db ATTEN
(100 VOLTS/DIV)
5 KHz



Ⓑ 60 db ATTEN
(100 VOLTS/DIV)
5 KHz
(RISE TIME)

- a) Output signal amplitude:
0 TO 7100 VOLTS
- b) Pulse width:
2 TO 300 NSEC
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
O.K.
- g) Prime power:

120 / 240V
50 - 60 Hz
[Signature]