


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-3-PS-N-PW-M*

S.N.: *3486*

Date: *JAN 27 87*



*40 db ATTEN*  
*10 VOLTS/DIV*  
*1.0 MHz*

- a) Output signal amplitude:  
*0 TO ± 20 VOLTS*
- b) Pulse width:  
*5 TO 100 NSER.*
- c) Rise time:  
*≤ 300 PSER*
- d) Fall time:  
*≤ 300 PSER*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*