


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *MP-70-15-C-7*

S.N.: *3487*

Date: *NOV 25 86*



*(A) 40 db ATTEN  
(5 VOLTS/DIV)  
1.0 MHz*



*(B) 40 db ATTEN  
(5 VOLTS/DIV)  
1.0 MHz*

- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 nSEC.*
- c) Rise time:  
*≤ 60 PSEC*
- d) Fall time:  
*≤ 135 PSEC.*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz.*