


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-1A-C-P-PN-M*

S.N.: *3489*

Date: *DEC 16 '86*



40 dB ATTEN
5 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz

[Signature]