


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *MT-V-HB3-C-M-N*

S.N.: *3490*

Date: *JAN 7 87*



*60 db ATTEN  
100 VOLTS/DIV  
5 KHz*



*60 db ATTEN  
100 VOLTS/DIV  
5 KHz*

- a) Output signal amplitude:  
*0 to 100 V.*
- b) Pulse width:  
*2 to 100 nsec*
- c) Rise time:  
*≤ 200 psec*
- d) Fall time:  
*≤ 2 nsec*
- e) PRF:  
*0 to 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz*