


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-N*

S.N.: *3493*

Date: *MIKE 10 87*



60 dB ATTEN
100 VOLT/DIV
50 KHz

- a) Output signal amplitude:
0 TO - 250 VOLTS
- b) Pulse width:
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
250 mA (MAX)
[Signature]