


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-PW-1*

S.N.: *3498*

Date: *OCT 22 86*



- a) Output signal amplitude:
0 TO + 200 V.
- b) Pulse width:
50 NSER TO 1.0 USER.
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
≤ 10 NSER
- d) Fall time:
≤ 10 NSER
- e) PRF:
MIN PW: 0 TO 100 KHz
MAX PW: 0 TO 5 KHz
- f) Jitter, stability: *(0.5% MAX DUTY CYCLE)*
OK
- g) Prime power:

50 VOLTS/DIV.
200 NSER/DIV
R_L = 50 Ω
5 KHz.

+24 ± 1.0 VOLTS
550 mA
(MAX)