


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVP-AV-15-PN*

S.N.: *3505*

Date: *JAN 26 87*



*40 db ATTEN
5 VOLTS/DIV
1.0 MHz*



*40 db ATTEN
5 VOLTS/DIV
1.0 MHz*

- a) Output signal amplitude:
0 TO ± 5.0 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*± 15 ± 1.0 VOLTS
250 mA (MAX)*