


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMH-54-C-N-05*

S.N.: *3506*

Date: *JAN 19 87*



- a) Output signal amplitude:  
*0 TO -10 VOLTS.*
- b) Pulse width:  
*≤ 130 PSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*  
*[Signature]*

*40 db ATTN*  
*10 VOLTS/DIV*  
*1.0 MHz.*