


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVMMN-1-V

S.N.: 3519

Date: JAN 9 87



Ⓐ 40 db ATTEN
5 VOLTS/DIV
7 MHz



Ⓑ AS Ⓐ BUT
50 MHz

- a) Output signal amplitude:
0 TO -5V
 - b) Pulse width:
0.2 TO 6 NS OR
 - c) Rise time:
≤ 100 PSEC
 - d) Fall time:
≤ 200 PSEC
 - e) PRF:
0 TO 50 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
+24 ± 1.0 VOLTS
400 mA (MAX)
- [Signature]*