



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMAN-1-N*
S.N.: *3521*
Date: *JAN 9 '87*



① 40db ATTEN
0.5 VOLTS/DIV
17 MHz



② 1ns (A) BUT
50 MHz

- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0.2 TO 6.0 nSEC
- c) Rise time:
6 100 PSEC
- d) Fall time:
6 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 VOLTS
400 mA (MAX)

[Signature]