


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVP-AV-15-C-P*

S.N.: *3522*

Date: *NOV 27 '86*



*40 db ATTEN
(5 VOLTS/DIV)
1.0 MHz*



*40 db ATTEN
(5 VOLTS/DIV)
1.0 MHz*

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240V
50-60 Hz*