


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMMA-2-P-6A-C*

S.N.: *3523*

Date: *DEC 17 '86*



40 db ATTEN
10 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO + 20 VOLTS
- b) Pulse width:
20 TO 200 NSEC
- c) Rise time:
(20% MAX DUTY CYCLE)
 $\le 3\text{ NSEC}$
- d) Fall time:
 $\le 3\text{ NSEC}$
- e) PRF:
0 TO 10 MHz
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz