


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-1-PS-05-14A-P*

S.N.: *3535*

Date: *FEB 26 87*



R_o = 10K
50 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *TO 710K*
50 NSEC TO 1.0 USEC
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC
- e) PRF:
0 TO 200 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 W