


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-A-1-PS-05-10A-N*

S.N.: *3537*

Date: *FEB 26 87*



*R<sub>i</sub> = 10k  
50 kHz*

- a) Output signal amplitude:  
*0 TO -100 VOLTS*
- b) Pulse width: *7 10k*  
*50 NSEC TO 1.0 USEC*
- c) Rise time:  
*≤ 5 NSEC*
- d) Fall time:  
*≤ 5 NSEC*
- e) PRF:  
*0 TO 200 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240V  
50-60 Hz*

*[Signature]*