


PULSE GENERATOR

PERFORMANCE CHECK


Model: *APR-32-PS-PW-TR-05-10A7*

S.N.: *3538*

Date: *MAR 4 87*



*Ⓐ 50 db ATTEN
∴ 32 VOLTS/DIV
200 KHz
RL = 50 Ω*



*Ⓑ AS Ⓐ BUT
RISE TIME.*

- a) Output signal amplitude:
0 TO +50 VOLTS.
- b) Pulse width:
20 TO 120 NSEC
- c) Rise time:
≤ 500 PSEC
- d) Fall time:
≤ 5 NSEC
- e) PRF:
0 TO 200 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50-60 Hz*