


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV-107B-PS-P*

S.N.: *3545*

Date: *APRIL 30 '87*



*R<sub>L</sub> = 30  $\Omega$*

*$\approx 0.6$  AMP/DIV*

*20 KHz.*

- a) Output signal amplitude:  
*0 TO +2.0 AMPS*
- b) Pulse width:  
*20 TO 200 NSEC.*
- c) Rise time:  
 *$\leq 10$  NSEC*
- d) Fall time:  
 *$\leq 10$  NSEC*
- e) PRF:  
*0 TO 20 KHz.*
- f) Jitter, stability:  
*OK.*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*

*[Signature]*