


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-93-C-P-05-CA-EN*

S.N.: *3547*

Date: *DEC 18 '86*



*60 db atten
100 V/DIV
20 KHz.
R_L = 50 Ω*

- a) Output signal amplitude:
0 TO +100 V.
- b) Pulse width:
20 TO 200 nsec
- c) Rise time:
55 nsec
- d) Fall time:
55 nsec
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*