


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV4-MPT P*

S.N.: *3548*

Date: *JAN 19 87*



- a) Output signal amplitude:  
*≥ +30 VOLTS*
- b) Pulse width:  
*≤ 1.0 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 db ATTEN*

*32 VOLTS/DIV*

*1.0 MHz*

*+15 ± 1.0 VOLT*

*150 mA (MAX)*

*[Signature]*