


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *3549*

Date: *JAN 19 87*



- a) Output signal amplitude:
≥ +30 VOLTS
- b) Pulse width:
≤ 1.0 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 dB ATTEN

32 VOLTS/DIV

1.0 MHz

+15 ± 1.0 VOLT

150 mA (max)

[Signature]