


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM1-2-N-M-ET-EN-ED*

S.N.: *3555*

Date: *FEB 25 87*



40 dB ATTEN
10 VOLTS/DIV
10 MHz

- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLT
500 mA (MAX)