


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AN-1413-C-PN-OT-MA*
S.N.: *3564* *-EA-EN*
Date: *FEB 19 87*



50 Ω ATTEN
= 32 VOLTS/DIV
1.0 MHz

- a) Output signal amplitude:
0.70 ± 40 VOLTS
 - b) Pulse width:
0.4 TO 2.0 NS
 - c) Rise time:
≤ 150 psec
 - d) Fall time:
≤ 250 psec
 - e) PRF:
0 TO 1.0 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
120/240V
50-60 Hz
- [Signature]*