


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AP-M-2-C-M-M-CA-EW-ED-W*

S.N.: *3565*


Date: *DEC 11 66*



Ⓐ 40 db ATTN
∴ 10 VOLTS/DIV
50 KHZ



Ⓑ AS Ⓐ BUT
RISE TIME



Ⓒ AS Ⓑ BUT
FALL TIME

- a) Output signal amplitude:
0 TO ±10 VOLTS
- b) Pulse width:
20 TO 500 NSER.
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ.
- e) PRF:
0 TO 50 KHZ.
- f) Jitter, stability:
OK.
- g) Prime power:

120/240V

50-60 HZ