


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVL-2-PS-M-PN

S.N.: 3579

Date: FEB 4 87



- a) Output signal amplitude:
0 TO ± 350 VOLTS
- b) Pulse width:
0 TO 100 NSER
- c) Rise time:
 ≤ 1.5 NSER
- d) Fall time:
 ≤ 3.0 NSER
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

70 dB ATTEN
 ≈ 320 VOLTS / DIV
2 KHz

120 / 240 V
50 - 60 Hz