


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PI-V-141-C-PN*

S.N.: *3582*

Date: *NOV 21 '86*



*50 db ATTEN
20 KHz*



*50 db ATTEN
20 KHz*

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100 nSEC
- c) Rise time:
≤ 250 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/2.40V
50-60 Hz*

[Signature]