


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MMA-3-EN-P*

S.N.: *3587*

Date: *JAN 21 87*



Ⓐ 40 db ATTEN.
10 VOLTS/DIV



ⓓ 40 db ATTEN.
10 VOLTS/DIV

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 PER SEC
- f) Jitter, stability:
OK
- g) Prime power:
+24 TO +28 VOLTS
400 mA (MAX)

[Signature]