


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PCR-B2-C-W-PN*

S.N.: *3592*

Date: *MAR 4 '87*



*P_{out} R_L = 50 Ω
20 kHz*



*N_{out} R_L = 50 Ω
20 kHz*

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
0.1 TO 5.0 μ SEC
- c) Rise time:
 ≤ 5 NS
- d) Fall time:
 ≤ 5 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*