

PULSE GENERATOR
PERFORMANCE CHECK


Model: ANN-1-PS-ATT 2

S.N.: 3600

Date: Feb 12 87

- a) Output signal amplitude:
ECL LEVELS, $-0.8 \pm$
- b) Pulse width: -1.8 VOLTS.
- c) Rise time: 50% DUTY CYCLE
FIXED.
- d) Fall time: ≤ 3 NSEC
- e) PRF: ≤ 3 NSEC
- f) Jitter, stability: 10 MHz to 250 MHz.
- g) Prime power: OK.

A OUTPUT.
2.0 NSEC/DIV (≈ 16 MHz)
1.0 VOLT/DIV



120/240 V
50-60 Hz

[Handwritten signature]

B OUTPUT
2 NSEC/DIV (≈ 200 MHz)
1.0 VOLT/DIV