


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-07*

S.N.: *3602*

Date: *JAN 28 87*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*40 db ATTEN
10 VOLTS DIV
100 KHz*

*120/280V
50-60 Hz*

[Signature]