


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AK-UP*

S.N.: *3604*

Date: *DEC 23 86*



40 dB ATTEN
5 VOLTS/DIV
50 KHz

- a) Output signal amplitude:
> + 5 VOLTS
- b) Pulse width:
1 TO 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
250 mA (MAX)