


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-11P*
S.N.: *3608*
Date: *DEC 8 86*



- a) Output signal amplitude: *20 TO 100 AMPS*
- b) Pulse width: *25 NSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power:

≈ 100 VAMPS/DIV
10 NSEC/DIV
1 KHz

+15 ± 1.0 VOLT
250 mA (MAX)

NA