


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *3611*  
Date: *DEC 8 '86*



- a) Output signal amplitude: *20 TO 100 AMPS*
- b) Pulse width: *25 NSEC.*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*100 AMPS/PIU*  
*10 NSEC/PIU*  
*1 KHz*

*+15 ± 1.0 VOLT.*  
*250 mA*  
*(MAX)*  
*[Signature]*