


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APG-1-17*
S.N.: *3612*
Date: *DEC 8 86*

- a) Output signal amplitude:
20 TO 100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1KHz
- f) Jitter, stability:
OK
- g) Prime power:



100 AMPS/DIV
10 NSEC/DIV
1 KHz

+15 ± 1.0V
250 mA
(max)
[Signature]