


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P-C*

S.N.: *3618*

Date: *FEB 20 '87*



*≈ 60 AMPS/DIV*  
*1 KHz.*

- a) Output signal amplitude:  
*+30 TO 100 AMPS*
- b) Pulse width:  
*25 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK.*
- g) Prime power:  
*20/240V*  
*50-60Hz*

*NA*