


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-172-P  
S.N.: 3619  
Date: FEB 24 87



50 dB ATTEN  
32 VOLTS/DIV.  
1.0 MHz

- a) Output signal amplitude:  
7.1 ± 60 VOLTS
- b) Pulse width:  
6.2 NS
- c) Rise time:  
N/A
- d) Fall time:  
N/A
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
4135 ± 5 VOLTS

5.0 mA (MAX)