


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-M-HV3-C-05-N*  
S.N.: *3621*  
Date: *MAR 26 57*



*50 dB ATTEN*  
*≈ 32 VOLTS/DIV.*  
*1.0 MHz*

- a) Output signal amplitude:  
*0 TO -40V*
- b) Pulse width:  
*0.4 TO 2.0 nSEC*
- c) Rise time:  
*≤ 150 pSEC*
- d) Fall time:  
*≤ 250 nSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*  
*MF*