


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVN-1-C-PN*

S.N.: *3629*

Date: *MAR 25 87*



Ⓐ *100 MHz.
40 dB ATTEN
5 VOLTS/DIV*



Ⓑ *AS Ⓐ BUT
250 MHz.*

- a) Output signal amplitude:
0 TO ±5 VOLTS.
- b) Pulse width:
0.2 TO 1.0 NSEC.
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*

[Signature]