


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PH-121-P*

S.N.: *3632*

Date: *77 APR 3 87*



*60 db ATTEN*  
*50 kHz*

- a) Output signal amplitude: *0 TO +100 VOLTS*
  - b) Pulse width: *≤ 1.0 NSGZ*
  - c) Rise time: *NA*
  - d) Fall time: *NA*
  - e) PRF: *0 TO 100 kHz*
  - f) Jitter, stability: *OK*
  - g) Prime power: *+15 ± 1.0 VOLTS*  
*150 mA (MAX)*
- [Signature]*