


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-AV-1-C-17

S.N.: 3633

Date: FEB 5 '87



- a) Output signal amplitude:
0 TO + 100 VOLTS
- b) Pulse width:
3 TO 100 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN

100 VOLTS/DIV

5 KHz.

$R_L = 50 \Omega$

120/240V

50-60 Hz