


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-100PT-1*

S.N.: *3637*

Date: *JAN 19 1987*



- a) Output signal amplitude: *7.5 ± 30 VOLTS*
- b) Pulse width: *5 1.0 NSEC*
- c) Rise time: *N/A*
- d) Fall time: *N/A*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *15 ± 1.0 VOLTS*
150 mA (MAX)

50 dB
32 VOLTS/DIV
1.0 MHz

[Signature]