


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *3638*

Date: *JAN 19 87*



a) Output signal amplitude:

≥ +30 VOLTS

b) Pulse width:

≤ 100 NS

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.00 VOLTS

150 mA (MAX)

[Signature]

50 dB

32 VOLTS/DIV

100 MHz