


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANIM-3-C-M-P-PN-0T*

S.N.: *3640*

Date: *MAY 7 87*



*40 dB ATTEN*  
*10 VOLTS/DIV*  
*12.0 MHz*

- a) Output signal amplitude:  
*0 TO ±15 VOLTS*
- b) Pulse width:  
*2 TO 15 NS*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 MHz*

*[Signature]*