


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMN-1-N.*

S.N.: *3659*

Date: *MAR 18/87*



40 dB ATTEN

5 VOLT/DIV

40 MHz PRF

- a) Output signal amplitude:
0 TO -5 VOLTS
- b) Pulse width:
0.2 TO 8.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 VOLT

500 mA (MAX)