


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMN-1-N*

S.N.: *3661*

Date: *MAR 18/87*



*40 db ATTEN*  
*.5 VOLTS/DIV*  
*40 MHz PRF*

- a) Output signal amplitude:  
*0 TO -5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+ 24 ± 1.0 VOLT*  
*500 mA (MAX)*