


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMN-1-N*

S.N.: *3662*

Date: *MAR 18/87*



40 db ATTEN
∴ 5 VOLTS / DIV
40 MHz PRF

- a) Output signal amplitude:
0 TO -5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLT
500 mA (MAX)